



SCK0F26-3.5 Mechanical Calibration Kit

Datasheet

Saluki Technology Inc.

Preface

Thanks for choosing Saluki Technology Inc Product.

We devote ourselves to meeting your demands, providing you high-quality measuring instrument and the best after-sales service. We persist with "superior quality and considerate service", and are committed to offering satisfactory products and service for our clients.

Document No.

SCK-02-05

Version

Rev03 2017.09

Saluki Technology

Document Authorization

The information contained in this Document is subject to change without notice. The power to interpret the contents of and terms used in this Document rests with Saluki.

Saluki Tech owns the copyright of this Document which should not be modified or tampered by any organization or individual, or reproduced or transmitted for the purpose of making profit without its prior permission, otherwise Saluki will reserve the right to investigate and affix legal liability of infringement.

Product Quality Assurance

The warranty period of the product is 36 months from the date of delivery. The instrument manufacturer will repair or replace damaged parts according to the actual situation within the warranty period. The user should return the product to the manufacturer and prepay mailing costs. The manufacturer will return the product and such costs to the user after maintenance.

Product Quality Certificate

The product meets the indicator requirements of the Document at the time of delivery. Calibration and measurement are completed by the measuring organization with qualifications specified by the state, and relevant data are provided for reference.

Quality/Environment Management

Research, development, manufacturing and testing of the product comply with the requirements of the quality and environmental management system.

Contacts

Service Tel: 886.2.2175 2930

Website: www.salukitec.com

Email: info@salukitec.com

Address: No. 367 Fuxing N Road, Taipei 105, Taiwan (R.O.C.)

1. Specification of SCK0F26-3.5 Calibration Kit

Model	Name	Frequency	Spec.	List	qty.
SCK0F26-3.5	3.5mm calibration kit (Calibration Level)	DC -26.5GHz	Open: $\pm 1.5^\circ$ Short: $\pm 1.5^\circ$ Load: VSWR<1.055 Adaptor: VSWR<1.106	Open: 3.5mm (M)	1
				Open: 3.5mm (F)	1
				Short: 3.5mm (M)	1
				Short: 3.5mm (F)	1
				Load: 3.5mm (M)	1
				Load: 3.5mm (F)	1
				3.5mm(f)-3.5mm(f)adaptor	1
				3.5mm(m)-3.5mm(m)adaptor	1
				3.5mm(m)-3.5mm(f)adaptor	1
3.5mm Type N torque wrench	1				

2. Specification of SCK0FE26-3.5 Calibration Kit

Model	Name	Frequency	Spec.	List	qty.
SCK0FE26-3.5	3.5mm calibration kit (Factory Level)	DC -26.5GHz	Open: $\pm 2^\circ$ Short: $\pm 2^\circ$ Load: VSWR<1.065 Adaptor: VSWR<1.12	Open: 3.5mm (M)	1
				Open: 3.5mm (F)	1
				Short: 3.5mm (M)	1
				Short: 3.5mm (F)	1
				Load: 3.5mm (M)	1
				Load: 3.5mm (F)	1
				3.5mm(f)-3.5mm(f)adaptor	1
				3.5mm(m)-3.5mm(m)adaptor	1
				3.5mm(f)-3.5mm(m)adaptor	1
3.5mm Type N torque wrench	1				

3. Specification of SCK0T26-M / F Calibration Kit

Model	Name	Frequency	Spec.	List	qty.
SCK0T26-M	3.5mm calibration kit Type: T (Male)	DC -26.5GHz	Open: $\pm 2.0^\circ$ Short: $\pm 2.0^\circ$ Load: VSWR<1.065	3.5mm (M) T type calibration kit	1
SCK0T26-F	3.5mm calibration kit Type: T (Female)	DC -26.5GHz	Open: $\pm 2.0^\circ$ Short: $\pm 2.0^\circ$ Load: VSWR<1.065	3.5mm (F) T type calibration kit	1

-END OF DOCUMENT-